

CORRIGENDUM NOTICE
AMENDMENT – II

RASHTRASANT TUKDOJI REGIONAL CANCER HOSPITAL AND RESEARCH CENTRE Tukdoji Square, Manewada Road, Nagpur – 440027. Email: admin@rstrech.com Phone No: 0712 – 2744441, 2748995			
Reference to tender no. R.S.T:R.C.H:2018/002. Below mentioned changes are made in tender.			
Sr. No.	Existing Tender Specification	Discussion / queries	Finalized
1.	Date of tender submission – 19.03.2018	Extension till 31.03.2018	Last date of submission is 26.03.2018
2.	Suggested payment against supply in USD	Payment terms in INR	Payment shall be in INR/USD For comparison of Rates values in INR shall be calculated at conversion at prevalent Rate
3.	EMD	Request in the form of Bank Guarantee	EMD shall be payable by DD / RTGS
4.	Payment terms not defined	Bidders wanted Payment terms	10% advance with order 70% by letter of credit 20% after satisfactory installation Testing and Trials of the machine
5	Delivery Schedule not mentioned	4 – 6 months by Elekta & Varian 6 – 8 months by Panacea	Delivery period may be 4-6 months
6	Offers are invited with terms & taxes & duties included	Bidders suggested that custom & taxes shall be extra as per standard terms of supply	Accepted
7	Penalty for downtime machine under breakdowns is Rs. 25000/- per day.	Bidders suggested that they proposes extended warrantee in ration of 1:2 or extra down time period instead of financial burden	Accepted
8	Permission and Licenses	Bidders wanted permission and licenses for operation of machine and to be made by users and extended support by suppliers	Purchaser is responsible to take all permission from regulatory authority AERB, and vendor will help in arranging the process.
9	Clause No. 5. Sr. No. 1	Photon Energy shall be 6X, 6MV and 15MV	6X, 10X, 6MV, 15MV
10	Clause No. 4.2.3	KV based SRS Capacity is required	KV based SRS Capacity is required
11	Clause No. 5 Sr. No. 16 MLC Thickness – 3mm / 5mm	2.5 mm OR 5 mm	2.5 mm OR 5.0 mm
12	Clause No. 5 Sr. No. 13	Beam Matching	Should be provided by vendor in machine standard format.
13	Clause No. 5 Sr. No. 20	Networking	PACS Delete
14	Clause No. 5 Sr. No. 20	Antivirus should be provided	Shall be in the scope of purchaser
15	Clause No. 5 Sr. No. 23	QA tools should be included	QA tools shall be arranged by purchaser.